


**SIMICON** 

**Quality certificate**

**Biological indicators SIMICON ST/DA-V6**

Biological indicators in compliance with ISO 11138-3 for steam sterilization at 121 °C/134 °C.

Organism: Geobacillus stearothermophilus  
spores ATCC 7953

Mean population (cfu): 1,6 × 10<sup>6</sup>  
determined by using agar surface method (TSA-Oxoid)

Carrier: filter paper

Packaging: paper / foil

Lot: V210724

Expiration date: 08.2026

D<sub>121°C</sub> – Value: 2,3 min

z-value: 7,6 °C

Survival time: D-value x (log cfu – 2)

Kill time: D-value x (log cfu + 4)

All data are determined at time of manufacturing.  
Resistance data all determined according to ISO 11138 and are only reproducible when exposed and cultured under the exact conditions used to obtain results report here. Results may vary under different test conditions.

Incubation: 7 days at 56 °C ± 2 °C in CSB

Storage:  
**at + 18 °C to + 25 °C and rel. humidity from 35 % to 70 %**  
Protect from light and sterilants.

Do not use after expiration date. Sterilize all cultures before disposal.

Date: 26.08.2024 Release: P.Barone

SIMICON GMBH • Sigmund-Riefler-Bogen 19 • D - 81829 München  
Tel.: +49 / (0)89 / 67 33 66 0 • Fax: +49 / (0)89 / 67 33 66 22

2024.10.16

TDG

26.08.2024

---

Date



Hygiene & Mikrobiologie

SIMICON GMBH  
 Sigmund-Riefler-Bogen 19  
 81829 München  
 Tel.: +49-89-673366-0  
 Fax: +49-89-673366-22

---

Stamp / Signature