Certificate of Analysis

<u>Client</u>

E&O Laboratories Ltd Burnhouse Bonnybridge Scotland FK4 2HH

Sample:	PP1030 - Plate Count Agar (Standard)
Batch Number:	04172303
Expiry Date:	2024-12-17
Date Received:	2024-09-25
Date Tested:	2024-09-25
Date of Issue:	2024-09-30
Sample Condition:	Satisfactory
Sample Number:	474894



Burnhouse, Bonnybridge Scotland, FK4 2HH Telephone: 01324 840404 Fax:01324 841314 Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / ≥ 50% for selective media. Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Bacillus subtilis NCTC 10400	93	Cream colonies	Cream colonies
Escherichia coli NCTC 12241	90	Cream colonies	Cream colonies
Staphylococcus aureus NCTC 12981	90	White colonies	White colonies

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
	ing is performed in I ≤ 7 units dependi		99*. The inspection level is \geq 0.4% of the batch and
рН	6.9	7.0 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	2-8 - 6-8 Pale Straw	ED/SOP/009 by visual observation. Range measured using Pantone [®] 4 colour process guide

Certificate of Analysis

<u>Client</u>

E&O Laboratories Ltd Burnhouse Bonnybridge Scotland FK4 2HH

Sample:	PP1030 - Plate Count Agar (Standard)	Burnhouse, Bonnybridge Scotland, FK4 2HH
Batch Number:	04172303	Telephone: 01324 840404 Fax:01324 841314
Expiry Date:	2024-12-17	Email: info@eolabs.com
Date Received:	2024-09-25	
Date Tested:	2024-09-25	
Date of Issue:	2024-09-30	
Sample Condition:	Satisfactory	
Sample Number:	474894	

$\overline{\ }$	
E&O Labora	atories Ltd

Moisture	Conforms	95% +/- 3%	ED/SOP/053 with drying and gravimetri determination
		ate of analysis relate only	

18-22ml

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications *Sterility sampling is outwith scope of accreditation.



Fill Quantity

Conforms

Douglas Cameron Technical Manager, E&O Laboratories Ltd

ED/SOP/054 by gravimetric determination