

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** PP1051 - Pseud. Agar w Glycerol & CFC  
**Batch Number:** 04172581  
**Expiry Date:** 2025-01-09  
**Date Received:** 2024-10-11  
**Date Tested:** 2024-10-11  
**Date of Issue:** 2024-10-17  
**Sample Condition:** Satisfactory  
**Sample Number:** 475463

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media /  $\geq 50\%$  for selective media.

Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Pseudomonas fluorescens NCTC 10038	111	Green colonies	Green colonies
Burkholderia cepacia NCTC 10661	65	Growth/translucent grey	Growth/translucent grey

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241	Inhibited	Inhibited

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



E&O Laboratories Ltd

**Sample:** PP1051 - Pseud. Agar w Glycerol & CFC  
**Batch Number:** 04172581  
**Expiry Date:** 2025-01-09  
**Date Received:** 2024-10-11  
**Date Tested:** 2024-10-11  
**Date of Issue:** 2024-10-17  
**Sample Condition:** Satisfactory  
**Sample Number:** 475463

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size.			
pH	7.2	7.1 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	2-8 - 6-8 Pale Straw	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	18-22ml	ED/SOP/054 by gravimetric determination
Moisture	Conforms	95% +/- 3%	ED/SOP/053 with drying and gravimetric determination

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**E&O Laboratories Ltd**

**Sample:** PP1051 - Pseud. Agar w Glycerol & CFC  
**Batch Number:** 04172581  
**Expiry Date:** 2025-01-09  
**Date Received:** 2024-10-11  
**Date Tested:** 2024-10-11  
**Date of Issue:** 2024-10-17  
**Sample Condition:** Satisfactory  
**Sample Number:** 475463

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

\*Sterility sampling is outwith scope of accreditation.



1894

**Douglas Cameron**  
Technical Manager, E&O Laboratories Ltd