

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : DD018	Material Name : Oxidase Discs (50 discs / vial)	Lot No : 0000666349
Report No.: 40001515613	Date of Release & Report : 2024-09-26	Expiry Date : 2025-09

Appearance

Filter paper discs of 10 mm diameter

Cultural response

Typical oxidase reaction given by 18-48 hours old grown culture & observed within 5-10 seconds at 25-30°C.

Organism	Reaction Observed
Cultural response	
<i>Pseudomonas aeruginosa</i> ATCC 27853	positive : deep purplish blue colouration of disc
<i>Neisseria gonorrhoeae</i> ATCC 19424	positive : deep purplish blue colouration of disc
<i>Escherichia coli</i> ATCC 25922	negative : purplish blue colouration after 10 sec / no colour change
<i>Staphylococcus aureus</i> ATCC 25923	negative: purplish blue colouration after 10 sec / no colour change

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

Control Media :

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.

- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 and WHO GMP

- . The Quality Assurance Parameters are as per the guidelines specified in CLSI (NCCLS) document M22-A3 wherever applicable.

Storage & Shelf Life

Store between 2 - 8°C. Use before expiry date on the label.


STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.


This document has been produced electronically and is valid


Shraddha Raval

**Microbiologist/Sr.Executive
Microbiologist**


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager

2024-09-26