

**PRODUCT CODE and  
DESCRIPTION:**

WR086-03-0019 LOOP CLEANSING SOLUTION

LOT #	EXPIRATION DATE	QUANTITY
2503081	2027/02/25	3.000,00


Visual appearance verification according to internal SOP: passed

All reported results were obtained at time of Lot release.

COPAN WASP S.r.l. certifies that product quality requirements have been met, information stated above is correct, and that the Manufacturer Quality Management System is compliant to the standards EN ISO 13485:2016 Medical devices - Quality management systems – Requirements for regulatory purposes and EN ISO 9001:2015 Quality Management System – Requirements and following amendments.

2025/03/14

Certificate Date of Issue  
(yyyy/mm/dd)



Quality Assurance  
COPAN WASP S.r.l.  
Via A. Grandi, 32  
25125 – Brescia  
Italy

Tel. +39 030 2687218 | fax +39 030 5541880 | wasp@copangroup.com