

### Quality certificate

#### Biological indicators DH/HL

Biological indicator in compliance with ISO 11138-4 for dry heat sterilization processes.

Organism:	Bacillus subtilis spores
Mean population (cfu): determined by using agar surface method (TSA-Oxoid)	$2,5 \times 10^6$
Carrier material:	filter paper
Package:	medical paper
Lot:	1020824
Expiration date:	09.2026
D <sub>160 °C</sub> -value:	2,2 min
z-value:	50,6 °C
Survival time:	D-value x (log cfu - 2)
Kill time:	D-value x (log cfu + 4)

All data are determined at time of manufacturing.

Resistance data all determined according to ISO 11138 and are only reproducible when exposed and cultured under the exact conditions used to obtain results report here. Results may vary under different test conditions.

Storage:

**at + 18 °C to + 25 °C and rel. humidity from 35 % to 70 %**

Protect from light and sterilants.

Do not use after expiration date. Sterilize all cultures before disposal.

Date: 27.11.2024

Release: P. Barone

27.11.2024

Date



Stamp / Signature

OK  
20250131 AFA