

 S GR11.6 DOC 198 REV04 OF 02/05/2024	CERTIFICATE OF QUALITY ASSURANCE	N° 25- CRT01514
--	---	--------------------------------

CODE	LOT.N°	EXP. DATE
4504C	4504C	2501744
		03/02/2028

DEVICE DESCRIPTION:	REGULAR SIZE TIP IN 109MM LONG DRY TUBE
----------------------------	---

Goods have been EO treated and are guaranteed free of amplifiable human DNA, detectable DNase and RNase.

Goods have been manufactured according to the ISO 9001:2015 and to the ISO 18385:2016.

DNase test: swabs are extracted, and eluates are inoculated with DNA Ladder. The DNA Ladder must be intact to verify the absence of detectable DNase.

RNase test: swabs are extracted, and eluates are inoculated with RNA Ladder. The RNA Ladder must be intact to verify the absence of detectable RNase.

Human DNA contamination test: swabs are extracted and a portion of the extract is tested for absence of detectable human DNA with STR typing techniques.

All reported results were obtained following internal SOP instructions.

Copan Italia SpA certifies that all product quality requirements have been met and that all information stated above is correct.

24/03/2025

Certificate Date of Issue


Quality Representative
Copan Italia SpA
Via F.Perotti, 10 – 25125 Brescia (Italy)