Liofilchem® Certificate of Analysis

Page 1 of 1

ProductBatchExpiration dateViolet Red Bile Glucose Agar0219255022029.02.12

Ref. 610059 - 620059 - 6100595

Physical quality control	Specification	Results
Expected pH-value (25°C)	7.4 ± 0.2	7.4
Appearance of powder	Free-flowing, homogeneous	Conforms
Colour of powder	Beige to reddish-beige	Conforms
Appearance of prepared medium	Slightly opalescent	Conforms
Colour of prepared medium	Reddish-purple	Conforms

Microbiological Performance

Tested according to CLSI M22-A3, EN ISO 11133, Harmonized USP/EP/JP

Reference media: Tryptic Soy Agar

Productivity, Method of control: Quantitative

Inoculum: 50-100 CFU

Control strains	Incubation	Expected Results	Specification	Results
Escherichia coli WDCM 00012	Aerobic, $24 \pm 2 h$ 37 ± 1 °C	$P_R \ge 20.5$	Good growth, pink to red colonies with or without precipitation halo	Conforms
Escherichia coli ATCC® 8739	Aerobic, 18-24 h 30-35°C	$P_R \ge 20.5$	Good growth, pink to red colonies with or without precipitation halo	Conforms
Salmonella Typhimurium WDCM 00031	Aerobic, $24 \pm 2 h$ 37 ± 1 °C	$P_R \ge 20.5$	Good growth, pink to red colonies with or without precipitation halo	Conforms
Salmonella Enteritidis WDCM 00030	Aerobic, $24 \pm 2 h$ 37 ± 1 °C	$P_R \ge 20.5$	Good growth, pink to red colonies with or without precipitation halo	Conforms
Pseudomonas aeruginosa ATCC® 9027	Aerobic, 18-24 h 30-35°C	P _R ≥≥ 0.5	Good growth, pink to red colonies with or without precipitation halo	Conforms

Selectivity, Method of control: Qualitative

Inoculum: 104-106 CFU

Control strains	Incubation	Specification	Results
Enterococcus faecalis WDCM 00009	Aerobic, 24 ± 2 h 37 ± 1°C	Inhibition	Conforms

Batch Release

Approved

Date 21.02.2025 Signature Quality Control (This document has been established electronically and is valid without signature)